Instruction Manual • November 2005



SIEMENS

Safety Guidelines: Warning notices must be observed to ensure personal safety as well as that of others, and to protect the product and the connected equipment. These warning notices are accompanied by a clarification of the level of caution to be observed.

Qualified Personnel: This device/system may only be set up and operated in conjunction with this manual. Qualified personnel are only authorized to install and operate this equipment in accordance with established safety practices and standards.

Unit Repair and Excluded Liability:

- The user is responsible for all changes and repairs made to the device by the user or the user's agent.
- All new components are to be provided by Siemens Milltronics Process Instruments Inc.
- Restrict repair to faulty components only.
- Do not reuse faulty components.

Warning: This product can only function properly and safely if it is correctly transported, stored, installed, set up, operated, and maintained.

Note: Always use product in accordance with specifications.

Copyright Siemens Milltronics Process Instruments Inc. 2005. All Rights Reserved	Disclaimer of Liability
This document is available in bound version and in electronic version. We encourage users to purchase authorized bound manuals, or to view electronic versions as designed and authored by Siemens Milltronics Process Instruments Inc. Siemens Milltronics Process Instruments Inc. will not be responsible for the contents of partial or whole reproductions of either bound or electronic versions.	While we have verified the contents of this manual for agreement with the instrumentation described, variations remain possible. Thus we cannot guarantee full agreement. The contents of this manual are regularly reviewed and corrections are included in subsequent editions. We welcome all suggestions for improvement.
	Technical data subject to change.

MILLTRONICS® is a registered trademark of Siemens Milltronics Process Instruments Inc.

Contact SMPI Technical Publications at the following address:

Technical Publications Siemens Milltronics Process Instruments Inc. 1954 Technology Drive, P.O. Box 4225 Peterborough, Ontario, Canada, K9J 7B1 Email: **techpubs.smpi@siemens.com**

- For a selection of Siemens Milltronics level measurement manuals, go to: www.siemens.com/processautomation. Under Process Instrumentation, select *Level Measurement* and then go to the manual archive listed under the product family.
- For a selection of Siemens Milltronics weighing manuals, go to: www.siemens.com/processautomation. Under Weighing Technology, select *Continuous Weighing Systems* and then go to the manual archive listed under the product family.

TABLE OF CONTENTS

TITLE PAGE ABOUT THIS ... About This Manual 7 About the OCM-3 7 SPECIFICATIONS 9 Programmer 10 Transducer 11 **Temperature Sensor** 11 Cabling 11 **Communication Software** 11

=

INSTALLATION

Installing the OCM-3	13
Outline and Mounting	13
OCM-3 Layout	14
System Diagram	15
Installing the Transducer	16
Installing the Temperature Sensor	16
mA Output	17
Relays	17
Synchronization	18
Power Connections	19
Installing the Memory Back-up Battery	20
Communicating Via Computer	20
Installing the Programmer	20

START UP

General	21
Keypad	21
Legend	22
Initial Start Up	22
Fundamental Checks	25

OPERATION

Memory	27
Security	27
Units	27
Flow Calculation	28
Display	28
Damping	29
Relays	30
mA Output	30
Fail-Safe	31
Flow rate and Totalizing	31
Logging	32
Blanking	34
Temperature	34
Time and Date	34
Emulation Mode	35
Reset	35
Flow Velocity Input	36
Auxiliary Head Input	36
DC Output	37
Diagnostic Aids	37

'D' PARAMETER LISTING	39
'F' PARAMETER LISTING	41
'P' PARAMETER LISTING	43
'U' PARAMETERS FOR P3 PRIMARY ELEMENT	51
Simple Exponential Devices, P3 = 0	53
BS-3680 Rectangular Flume, P3 = 1	58
BS-3680 Round Nose Horizontal Crest Weir, P3 = 2	60
BS-3680 Trapezoidal Flume, P3 = 3	62
BS-3680 U - Flume, P3 = 4	64
BS-3680 Finite Crest Weir, P3 = 5	66
BS-3680 Thin Plate Rectangular Weir, P3 = 6	68
BS-3680 Thin Plate V-Notch Weir, P3 = 7	70
Rectangular Weir (Contracted), P3 = 8	72
Round Pipe, P3 = 9	74
Palmer-Bowlus Flume, P3 = 10	76
H - Flume, P3 = 11	78
Universal Head vs. Flow, P3 =12	80
Rectangular Area x Velocity, P3 = 13	82
Trapezoidal Area x Velocity, P3 =14	84
Modified Trapezoidal Area x Velocity, P3 = 15	86
U Channel Area x Velocity, P3 = 16	88
Circular Area x Velocity, P3 = 17	90
Gull-Wing Area x Velocity, P3 = 18	92
Egg-Shaped Area x Velocity, P3 =19	94
Universal Area x Velocity, P3 = 20	96

APPENDICES

Maintenance	99
Error Codes	100
Communications	101



ABOUT THIS ...

ABOUT THIS MANUAL

Although the OCM-3 is very 'approachable' due its dialogue capabilities and intuitive operation, the user should be familiar with this manual. This manual provides the user with the necessary information required to install, start up and operate the OCM-3.

As the OCM-3 prompts the user with specific messages in a step-by-step fashion during programming, the Start Up section serves essentially to compliment the OCM-3. Start Up provides the user with instructions on the use of the programmer and an overview of the programming requirements.

The 'D', 'F', 'P' and 'U' parameters listed in the Parameters section provide a quick reference of the available programming and display parameters and their options. The 'U' parameter listing also provides mathematical and graphical details as a reference to assist the user in programming the OCM-3 to the primary element being used. The user is urged to rely on the manufacturer's specification for obtaining and identifying the primary element to which the OCM-3 is being applied.

In short,

If you want to know about
the product
getting started
how it works

Read

About This . . . Specifications

Installation Start Up

Operation Parameters Appendices

ABOUT THE OCM-3

The OCM-3 is to be used only in the manner outlined in this instruction manual.

The Milltronics OCM-3, Open Channel Meter, is an electronic instrument designed to measure flow in open channels. It is housed in a polycarbonate enclosure and comes with a removable programmer. As a system, it is used in conjunction with a remote ultrasonic transducer (or auxiliary head measurement device) and a temperature sensor.

The OCM-3 transmits a pulse signal to the transducer which is then emitted as ultrasonic pulses. The pulses echo off the water surface and are then sensed by the transducer. The time for a pulse to echo back from the water surface is temperature compensated and converted into a measurement of head. The OCM-3 converts the head measurement into flow rate, but also provides a velocity sensor input for applications where a flow velocity measurement is required to perform the flow calculation. The flow rate is totalized and stored in a comprehensive data log to provide detailed flow analysis.

Programming of the OCM-3 allows the operator to select the flow calculation specific to the primary measuring device (flume, weir or pipe). Special emphasis has been placed on providing the most accurate flow calculations possible. To this end, specific routines have been written to comply with the British Standards Institute's Specifications BS-3680. These routines calculate correction factors taking into account second order effects such as approach velocity and boundary layer. In the event that flow measurement is not covered by one of the flow calculations provided, the OCM-3 can be programmed for flow measurement using one of the universal flow calculations.

The OCM-3 provides serial communication for remote programming, data log retrieval and print out for devices such as computers, PLCs and printers. Milltronics provides a standard utilities software package for OCM-3 programming, remote display and data retrieval. However, the user is not limited to the software provided. The user can develop his own software program to perform tasks suited to his specific needs.

The OCM-3 features:

- ✓ multi field illuminated LCD, for 'Flow and Total' and 'Relay Status' display
- ✓ 0 or 4 to 20 mA output
- ✓ three multipurpose relays, including remote totalization
- ✓ 1 to 24 months data log, subject to logging rate
- ✓ extensive serial communication, including RS-232
- ✓ removable infra-red programmer
- ✓ AC and DC (bi-current) operation.

SPECIFICATIONS ===

Power:	» dc supply: » ac supply:	» 9 to 30 V DC, 8 W r » 100/115/200/230 V 20 VA max	
Environmental:	 » location » altitude: » ambient tempera: » relative humidity » installation categories » pollution degree 	» suitable for out	
Memory back-up:	 » 3 V lithium batter » operating life 1 ye » 'SuperCap' capad during battery re 	citor for back-up	quivalent)
Range:	» 0.3 m min to 1.2 0.6 m min to 3 m	· · · ·	
Resolution:	» 0.2 mm (0.007")		
Accuracy:	» ± 1 mm/m, calculated error less than 0.02%		
Temperature Compensation:	» external sensor to compensate over the operating range		
Programming:	» via supplied prog communication li		
Inputs:	» velocity sensor a	nd	
	auxiliary head	» range: » resolution:	» 0 to 10 V dc » 2.7 mV
Outputs:	» transducer drive:		Ilses of 0.1 msec typical sec typical repetition rate.
	» instrumentation:		0-20 or 4-20 mA 5 uA

		» isolation:	300 V ac continuous
	» relays:	» 3 alarm/control relay	/S
			ntact per relay, rated at -inductive or 30 V dc
	» dc output:	» +24 V dc	
		» 20 mA average to 20 cycle max	00 mA at 1/10 duty
Communication:	» RS-232 or \pm 20 mA bipolar current loop,300, 600, 1200, 2400, 4800, 9600 or 19200 baud		
Data Logs:	» variable rate on 1	, 5, 15, 30 or 60 min or	[.] 24 hr
	» 31 days minimum	n/2 years maximum	
Display:	» illuminated liquid crystal 5 x 7 dot matrix display with 2 lines of 40 characters each		
Enclosure:	»Type 4X / NEMA 4X / IP65		
	» 209 mm W x 285 mm H x 92 mm D (8.2" W x 11.2" H x 3.6" D)		
	» polycarbonate		
Weight:	» 2.3 Kg (5.1 lb)		
Approvals:	» CE *, FM, CSA _N	RTL/C	
	» MCERTS Class 1	open channel flow dev	vice with environment
	operation limits at 3	85 °C (95 °F) at 93% re	lative humidity



SIRA MC 050058/01

PROGRAMMER

Enclosure:	» general purpose	
	» 67 mm W x 100 mm H x 25 mm D (2.6" W x 4" H x 1" D)	
	» ABS plastic	
Operating Temperature:	» −20 to 50 °C (−5 to 122 °F)	
7ML19985AB01		

Battery:		» 9 V (ANSI / NEDA 1604, PP3 or equivalent) or 3V lithium battery
TRANSDUCE	2	
Model:	» XRS-5*	
	Refer to associated	d Transducer manual.
TEMPERATUR	RE SENSOR	
Model:	» TS-2, LTS-1 or I	_TS-1C
	Refer to associated Temperature Sensor manual.	
CABLING		
Transducer:		» RG-62U coaxial » maximum separation 183 m (600 ft)
mA Output:		 » must be run in grounded metal conduit » Belden 8760 or equivalent
Synchronisation:		» Belden 8760
Temperature Ser	nsor:	» Belden 8760, 1 pair shielded/twisted, 18 AWG or equivalent
		» maximum separation 183 m (600 ft)
		» can be run in conduit with transducer cable
Communication:	» RS-232:	» Belden 8770, 3 wire shielded, 24 AWG or equivalent
		» maximum separation 15 m (50 ft)
	» Bipolar Current:	» Belden 9552, 2 pair shielded/twisted, 18 AWG or equivalent
		» maximum separation 1,500 m (5,000 ft)
*Note: The XRS-5 must be used with the TS-2 external temperature sensor when operating with the OCM-3.		

Velocity Input:	» Belden 8760 or equivalent
-----------------	-----------------------------

Auxiliary Input:

» Belden 8760 or equivalent

24 V Output:

» Belden 8760

COMMUNICATION SOFTWARE

Milltronics Utilities Software on standard PC floppy disk for DOS 3.1 and up.



INSTALLATION

Installation shall only be performed by qualified personnel and in accordance with local governing regulations.

INSTALLING THE OCM-3

The OCM-3 should be mounted in a clean, dry area that is: within the ambient temperature range and suitable for the specified enclosure. The front cover should be accessible for programming and viewing.

It is advisable to keep the OCM-3 away from high voltage or current runs, contactors and SCR control drives.

Do not mount the OCM-3 in direct sunlight without the use of a sun shield.

This product is susceptible to electrostatic shock. Follow proper grounding procedures.

OUTLINE AND MOUNTING



Non metallic enclosure does not provide grounding between connections. Use grounding type bushings and jumpers.

OCM-3 LAYOUT



All field wiring must have insulation suitable for at least 250 V.

ackslash Hazardous voltage present on transducer terminals during operation.

 Δ dc terminals shall be supplied from SELV source in accordance with IEC 1010-1 Annex H.

Relay contact terminals are for use with equipment having no accessible live parts and wiring having insulation suitable for at least 250 V.

The maximum allowable working voltage between adjacent relay contacts shall be 250 V.



Maximum system capability. Not all components or their maximum quantity may be required.

INSTALLING THE TRANSDUCER



Basic Wiring – Transducer

Hazardous voltage present on transducer terminals during operation.

Note: When using the XRS-5 transducer with the OCM-3, use the TS-2 external temperature sensor. The internal temperature sensor in the XRS-5 cannot be used.

INSTALLING THE TEMPERATURE SENSOR



Basic Wiring – Temperature Sensor

In order to compensate for uniform temperature change in the air between the transducer and the flow surface, the temperature sensor must be connected to the OCM-3.

Maximum cable run 183 m (600 ft) of Beldon 8760, 1 pair shielded/twisted, 18 AWG or equivalent.

Temperature sensor cable can be run with the transducer cable in a grounded metal conduit.

Ground shield at OCM-3 only.

mA OUTPUT



isolated 0 or 4 to 20 mA output (P26) into 1 K Ω load maximum. Wiring should conform to standard instrumentation practices. Ground shield at OCM-3 only.

RELAYS



relays shown in de-energized state, contacts rated at 5 A at 250 V non-inductive.

All relays are certified for use in equipment where the short circuit capacity of the circuits in which they are connected is limited by fuses having ratings not exceeding the rating of the relays.

SYNCHRONIZATION

Where two to a maximum of twelve transducers will be sharing a common conduit the OCM-3s should be synchronized. In order to synchronize OCM-3s:

» remove jumper J1 on board A on all but one OCM-3



» interconnect the SYNC terminal (TB1-20) of all OCM-3s. Insure that all OCM-3s share a common ground (TB1-34).



POWER CONNECTIONS

The OCM-3 power supply accepts 100, 115, 200 or 230 V ac per switch SW1 (board B) selection and 9 to 30 V dc.

The OCM-3 operates either under ac or dc power, or both ac and dc live simultaneously. If both ac and dc power are live, the OCM-3 normally draws power from the ac supply. In the event that the ac supply fails, the OCM-3 then draws power from the dc supply.

AC POWER



* switch shown in 'OFF' position, select appropriate voltage.

select voltage via switch on board B

The equipment must be protected by a 15 A fuse or circuit breaker in the building installation.

A circuit breaker or switch in the building installation, marked as the disconnect switch, shall be in close proximity to the equipment and within easy reach of the operator.





dc terminals shall be supplied from SELV source in accordance with IEC 1010-1 Annex H.

INSTALLING THE MEMORY BACK-UP BATTERY



Disconnect power before installing or replacing the battery.



Do not install the memory back-up battery until the OCM-3 is to be used.

The unit is supplied with one battery package. Remove the battery from the package and insert it into the battery socket.

Refer to Operation \ Memory.

The memory battery, B1 (see Specifications) should be replaced yearly to insure memory back up during lengthy power outages. An on board capacitor provides one hour of memory retention in order to preserve the memory while the battery is being changed.

COMMUNICATING VIA COMPUTER

Refer to Communication.

INSTALLING THE PROGRAMMER

To program the OCM-3 via the Programmer, it must be placed into the front cover recess of the OCM-3. The back of the Programmer has a magnetic plate which will hold the programmer in place. It can be removed when programming is completed.



START UP

For the initial start up, it is advisable to use the programmer for programming the OCM-3. The programmer transmits a coded infrared signal each time a key is pressed.

The OCM-3 is designed to automatically scroll through the 'A', 'D', 'F', 'P' and 'U' parameters in a structured sequence. The scrolling is interactive in that, depending on the option chosen for a given parameter, subsequent parameters may be skipped or modified. The user is thus prompted to satisfy only the parameters which are available to him for the application he has chosen.

KEYPAD



LEGEND

Press the associated programmer key:		
Display shown on OCM-3:	MILLTRONICS OCM III Copyright Milltronics, 1992	
Programmer key:		

INITIAL START UP

After installation procedures have been completed, the OCM-3 may be powered up. Upon initial powering up, the unit momentarily displays:



and then scrolls through the available languages:



The OCM-3 is asking which language you prefer to communicate in!





factory set security code 2.71828 must be entered



For optimum calibration accuracy, an 'F13' should be performed prior to accessing 'F2', the normal operating mode.



If data logging is desired, the time and date must be set.



The start up procedure is now complete. Enter 'F2' to place the OCM-3 in the normal operating mode.



Note: To save parameter values, return to RUN mode (F2) after programming.

The OCM-3 now displays the flow rate and total. Refer to Operation \ Display

FUNDAMENTAL CHECKS

For accurate determination of flowrate, accurate head measurement is essential. Check the following and correct if necessary.

- » check D5 for correct temperature at transducer location.
- » check D9 for correct distance from transducer to head.
- » check D0 for accurate head measurement.



OPERATION

Upon power up, the transducer is fired periodically as set by P36. A long interval between measurements may be desirable in order to conserve power* when operating the OCM-3 from a DC source of limited capacity.

The echo is processed to determine the head (D0). The flow rate (D1) is calculated by the OCM-3 as a mathematical function (P3 and P4) of head or a function of head and velocity (P42). The flow rate is then integrated to yield the totalized flow (D2). The 'Flow' and 'Total' fields which are displayed during the normal running mode (F2) are also continuously updated.

Viewing or changing the content of a parameter (except F1, emulation) is done without disturbing the acquisition, processing or logging of flow data (see \ Security).

MEMORY

During a power interruption, the memory back up will hold the programming, the log and the totalizer values, and run the clock. The memory battery (B1) provides up to one year of memory retention (see Appendices \ Maintenance).

Note: To save parameter values, return to RUN mode (F2) after programming.

SECURITY

The content of all 'A', 'D', 'F', 'P' and 'U' parameters can be viewed without having to satisfy the security parameter, F0. However if it is desired to change the content of any of these parameters, the security parameter must be satisfied (except for resetting the running min/max displays, parameters D3/D4 and D6/D7).

Once security has been satisfied, access continues for 5 minutes after the last key is pressed or until F2 is re-entered.

The security code may be changed from its factory set value, 2.71828, by entering a new value into F10. It is imperative that the new value be recorded, as the code can not be viewed. If the code is lost, consult Milltronics.

UNITS

Programming of the OCM-3 involves setting the units of measure:

- » P1 linear and velocity
- » P2 temperature
- » P5 flow rate and volume

If the units are changed during the course of operation, the change will be effected through all associated parameters and displays and will rescale flow and total data stored in the logs.

^{*} restricted usage of display lighting (P14), relays (P15, 18 & 21), mA output and communications will also conserve power.

FLOW CALCULATION

Absolute vs. ratiometric

The OCM-3 can be programmed to use either of two methods (P4) for calculating flow from the head measurement: absolute or ratiometric. The result is the same regardless of the method used. The principal difference is the information that must be entered in order for the OCM-3 to carry out the calculation. The user's choice of method may ultimately be based upon the information which is at hand. Refer to U parameters for the primary element selected for a listing of the information required.

For the ratiometric method, it is usually sufficient that the user know the flow rate (Q_{cal}) which occurs at maximum head (h_{cal}) .

On the other hand, absolute calculations require that the user enter information such as: the physical dimensions of the primary element and the constant relating to units of measure for both linear dimensions and flow rates.

e.g.

the general formula for flow through a single exponent primary element is:

 $Q = KH^{x}$

the specific formula for flow through a 45° V-notch weir is:

thus: Q = flow in cubic feet per second K = constant of 1.03 H = head in feet

The absolute method is not applicable to the following:

Palmer Bowlus flume H flume

DISPLAY

The normal display during operation is the Flow and Total Display (F2).







flashing) The OCM-3 provides illumination for the LCD for easier viewing of the display. Illumination can be set (P14) to be normally on or off, or automatic. When automatic

Illumination can be set (P14) to be normally on or off, or automatic. When automatic is selected, the lighting will automatically go on when keypad activity is sensed and then extinguish after 15 seconds of inactivity.

For battery operation, set display lighting to off or auto.

DAMPING

The OCM-3 provides two separate damping functions: reading and mA output. Zero or no damping allows fastest response while high or 100% provides the slowest response. The damping is usually set to provide a reliable response without sacrificing stability.

The reading damping, P13, dampens only the flow rate reading of the 'Flow and Total' display F2. The damping selections are: off, low, medium and high. Relay functions associated with flow rate respond to the dampened reading values.

mA output damping, P27, dampens the change in the mA output. The parameter entry is in seconds for spanning the 0 to 100% of the mA range selected (P26). Displays and relay functions associated with the mA output respond to its dampened value.

7ML19985AB01

RELAYS

Three on board multipurpose relays are provided by the OCM-3. P15, 18 and 21 set the respective functions for relays 1, 2 and 3. Depending on the function selected, these parameters determine the need and configuration of the subsequent relay control parameters, P16, 17 (relay 1); P19, 20 (relay 2) and P22, 23 (relay 3).

If the relay is to function as a driver for a remote totalizer or as a flow sampler contact, the totalizer multiplier (P32) will be factored by the setpoint . Note that parameters P16, P19, and P22 will default to zero. When a relay is set to totalizer (P32), you must have the applicable parameter (P16 for relay 1, P19 for relay 2, or P22 for relay 3) set to something other than zero (normally 1).

Example: For relay 1

Relay totalizer factor = <u>totalizer factor (P32)</u> relay 1 setpoint (P16)

> = <u>100 (P32=5)</u> 2 (P16)

= 50 units/pulse

The status of each relay is shown in the display. Refer to \ Flow and Total Display.



For battery operation, have relays energizing on alarm.

mA OUTPUT

The OCM-3 provides a mA output (TB1-4/5) which can be assigned (P24) to represent the measurement of flow, head, velocity or temperature. The associated scaling, P25, is factory set to a value of '0'. This provides normal scaling with respect to the assigned measurement.

Normal scaling for representation of flow, head or velocity is:

» 0 or 4 mA	= 0	
» 20 mA	= maximum measurement value for:	» P6 [*] : flow rate at maximum head
		» P7: maximum head
		» P10: velocity at maximum head

Normal scaling for representation of temperature is:

 $* 0 \text{ or } 4 \text{ mA} = -40 ^{\circ}\text{C}$

» 20 mA = 60 °C

If custom scaling is required, the 20 mA corresponding value (other than 0) can be entered into P25. The range (0 to 20 or 4 to 20 mA) and damping (see Damping) are set via P26 and P27 respectively.

The mA function can be overridden for test purposes by setting the desired mA value into F3. When the value is entered, the mA output will go to that value. When F3 is exited, the mA output will revert to normal operation. Also, see \ Emulation Mode.

*In the case of absolute calculations (P4=0), P6 is calculated by the OCM-3.

FAIL-SAFE

In the event of an echo loss, the fail-safe timer will begin counting. If the echo loss duration surpasses that of the time set (P29), a '*No Echo*' alert will be displayed in the Status field (see \ Display). The mA output will respond (P30) by either holding the last value or immediately going to a predetermined value (P31). The head and derived flow will hold their last value and totalization and logging will continue, based on that value. Upon resumption of a valid echo, the mA output will return to a value corresponding to the present value of the measurement assigned, at the mA damping rate (P27).

FLOW RATE AND TOTALIZING

Flow rate

Calculation of the flow rate is ongoing. It is normally viewed under the Flow and Total display (F2) with the decimal point set per P33. It can also be viewed under D1 as the raw flow calculation. Data on the running minimum and maximum flows that have occurred since the last reset can be viewed in two ways:

- » F7 gives the running min/max flows and their time and date of occurrence since the last reset. F7 is reset by F8 but only after satisfying the security parameter F0.
- » D3/D4 give the respective running min/max flow data, only, that have occurred since they were last reset. D3/D4 are reset simultaneously by entering 0 into either D3 or D4. D3 and D4 will then adopt the current flow rate and track the running min/max values from that point on. The security parameter (F0) does not need to be satisfied in order to reset D3/D4.

Flow data specific to a particular time and date can be viewed under the data log F14 (see \ Logging).

Totalizing

Totalizing of the calculated flow is ongoing. It is normally viewed under the Flow and Total display (F2).

An auxiliary totalizer (D2) is provided for operator usage and is intended for short term totalizing to a maximum count of 999999. It can be reset or preset independently of the F2 totalizer after satisfying the security parameter (F0).

In order to adjust the rate of filling of the totalizer, the totalizer multiplier (P32) can be set to an appropriate value. The totalizer can be reset via F11. Totalizing that is specific to the time and date can be viewed under the data log F14.

The OCM-3 can be programmed to operate a remote totalizer by assigning any of the relays (P15, 18 or 21) to act as a totalizer contact^{*}. Under this function, the maximum rate of contact closure is 2/sec with a closure duration of 200 msec^{**}.

* The totalizer count is set by the relay setpoint parameter (P16, 19 or 22 respectively).

** Typically the totalizer should be set for 300 to 3000 counts per day at maximum flow.

Under low flow conditions, a cut-off head (P45) can be entered to avoid totalizing flows occurring at or below the flow corresponding to the cut-off head.

LOGGING

The OCM-3 provides an extensive logging feature which can be viewed on the local display or retrieved via the serial communication link. The logging rate (P39) can be fixed or variable. The latter being useful in conserving logging space. The condition for variable logging is determined when selecting the logging rate.

Variable logging rate conditions are catagorized as : rate of change of flow, percent of maximum flow or percent of maximum head. Logging occurs at the normal (slower) rate while the condition is less than the setpoint (P40). If the condition exceeds the log rapid setpoint, the rapid rate of logging takes effect until the condition falls below the log normal setpoint (P41).

The associated setpoint units are : % change of maximum flowrate per minute, % of maximum flowrate and % of maximum head, respectively. The setpoints represent the absolute value of the rate of change; that is, for either increasing or decreasing flowrate. The OCM-3 does not recognize negative entries into P-40 and P-41.

Flow data is logged in 1/2 % increments from 0 to 110 % of maximum flow. Flows above 110 % are logged as 110 %. Truncation of flows to 110 % does not apply to daily totalization.

Log Capacity vs Rates

	<u>rate</u> 1 min	<u>capacity</u> 31 days
	5	3 months
	15	9 months
	30	1 year
	60	1.5 years
	24 hr	2 years
e.g.	15/5	9 months max / 3 months min

Once the log is filled, the old data will be successively written over with the new data being logged.

The log can be examined via F14. Viewing of the log is done by task and by method. The viewing tasks are: daily flow totals, flow rates and min-max flow data for flow or temperature. The viewing methods are: by first entry, by last entry and by specified date. The scrolling keys are used to maneuver through the tasks, methods and time of day.



The day totalizer (F14) does not use the master totalizer multiplier (P32). It is possible that the daily total overflows. In such a case the display will show +++++.

OCM III

BLANKING

Blanking is used to ignore the zone in front of the transducer where ringing or other false echo is at a level that interferes with the processing of the true echo. The minimum blanking is factory set, but can be overridden by entering the desired distance into P47.

Ringing is the inherent nature of the transducer mass to continue vibrating after the transducer has been fired. Ringing decays to acceptable levels in the order of milliseconds. Excessive cold or over tightening of the transducer mounting may increase the ring time such that it appears as an echo during the receive cycle. This is usually indicated by an incorrect high head reading. Excessive ring time may be overcome by increasing the blanking.

TEMPERATURE

The temperature as currently registered by the temperature sensor is viewed under D5. Data on the running minimum and maximum temperatures that have occurred since the last reset can be viewed in several ways:

F7 gives the running min/max temperatures and their time and date of occurrence since the last reset. F7 is reset by F8 but only after satisfying the security parameter F0.

D6/D7 give the respective running min/max temperature data, only, that have occurred since they were last reset. D6 is reset by entering a value lower than D5, and D7 is reset by entering a value larger than D5. D6 and D7 will then adopt the current temperature value and track the min/max values from that point on. The security parameter F0 does not need to be satisfied in order to reset D6 or D7.

Temperature data specific to the time and date can be viewed under the data log F14 (see \ Logging).

D14 indicates the resistance of the temperature sensor corresponding to the temperature shown in D5.

TIME AND DATE

If the data logging features of the OCM-3 are to be used, the time (F4) and date (F5) must be set. The day starts at 00:00:00 and ends at 23:59:59.

Adjusting the Time

If the clock time is advanced beyond the next anticipated logging time, the entry for each missed logging time is filled with a code which indicates that the system was not able to make entries at those times.

The daily total will be reduced proportional to the amount of time the day was shortened.

If the clock time is set back beyond the preceding logging time, the previously logged date will be written over with new data as the logging proceeds.

The daily total will be increased proportional to the amount of time the day was lengthened.

Adjusting the Date

If the calendar is reset, the OCM-3 will adjust the log dates accordingly, taking into account leap years and days per month.

EMULATION MODE

The flow calculation (P3/P4) can be checked for accuracy by using the emulation parameter F1. The head is entered and the corresponding flow is displayed. This function is useful when troubleshooting discrepancies between the OCM-3 calculation and the expected flow.

Relays assigned to functions associated with the emulation parameter respond to the emulated flows.

The mA output does not track the emulated flows when P28 (mA output emulation) = 0. However, if it is required to do so, then the emulator parameter should be set to 1.

RESET

The following resets can only be executed after satisfying the security access, F0.

Cold Start

If it is desired to reset all parameters, logs and totalizers to their factory setting, this is done by forcing a cold start, F12.

Master Totalizer

If it is desired to reset the master totalizer (F2), this is done by parameter F11.

Data Log

If it is desired to reset the data log (F14), this is done by parameter F15.

Min/Max Log

If it is desired to reset the min/max log (F7), this is done by parameter F8.

FLOW VELOCITY INPUT

In some applications, the flow calculation for the chosen primary element requires a velocity input. In this type of application, the transducer measurement is used to calculate the cross sectional area of the flow. By multiplying the area with the distance per time units of velocity, the volume per time units of flow are calculated. The calculated velocity can be viewed via D8.

The 0% and 100% limits of the velocity input must be scaled using parameters P8 and P9.

- » select P8
- » enter the voltage corresponding to zero velocity
- » select P9
- » enter the velocity corresponding to 5 V
- e.g. If the velocity sensor output is 1 V per m/sec and the output is scaled for 7 V at 100% velocity (7 m/sec), then enter 5 m/sec. If the output is scaled for 4 V at 100% velocity (4 m/sec), enter 5 m/sec.

P8 and P9 can only be accessed if P3 has been set for an option that requires the use of a velocity input. The input voltage level can be viewed via D12.

Voltage Input



Current Input



typical 1 - 5 V signal from velocity sensor.

typical 4 - 20 mA signal from velocity sensor. Add terminating resistor. e.g. 250 Ω for 1 - 5 V over 4 - 20 mA.

Signal must be positive with respect to ground.

Velocity Input (additional to Basic Wiring)

AUXILIARY HEAD INPUT

In some applications, the transducer input (TB1-1/2) is not used to provide a signal for head measurement. A typical example of this is an application which is beyond the 3 m (10 ft) range of the OCM-3. In such a case, the head could be derived from another Milltronics level monitor or other compatible device.

The method of head determination is set by P42. The OCM-3 simply substitutes the signal from the auxiliary device for the ultrasonic measurement provided by the transducer. The programming and consequent flow calculation are performed as normal.
The 0% and 100% limits of the auxiliary input must be scaled using parameters P43 and P44.

- » select P43
- » enter the voltage corresponding to zero head
- » select P44
- » enter the head corresponding to 5 V.
- e.g. if the head output is 1 V per m and the output is scaled for 7 V at 100% head (7 m), then enter 5 m. If the output is scaled for 4 V at 100% head (4 m), enter 5 m.

P43 and P44 can only be accessed if P42 has been set for head determination by an auxiliary device. The input voltage level can be viewed via D13.

Voltage Input 0 - 10V OUT S Н Y 2 M ٧ С Δ Е 0 U N 4 L D L М Х V M 16 17 18 19 I I 1

Current Input



typical 1 - 5 V signal from auxiliary device.

typical 4 - 20 mA signal from auxiliary device. Add terminating resistor. e.g. 250 Ω for 1 - 5 V over 4 - 20 mA.

Signal must be positive with respect to ground.

Auxiliary Input (additional to Basic Wiring)

DC OUTPUT

The OCM-3 provides a 24 V DC output that can be used to supply power to a remote customer device. The output is not isolated from the DC rails that supply the OCM-3 electronics and no overload protection is provided. As such, the output must not operate beyond its specified capacity.

DIAGNOSTIC AIDS

D15 through D18 are diagnostic aids to Milltronics service personnel when troubleshooting system problems.

D18 also serves as a performance indicator by displaying the number of valid echos received as a percentage of the number of pulses being fired by the transducer. A low value indicates that a large proportion of the pulses fired are not producing valid echos. In such instances, the transducer should be checked for proper mounting and aiming or the transducer wiring checked for excessive noise. If the value is zero, shorted or opened transducer wiring may be suspect.

'D' PARAMETER LISTING

Refer to 'Operation' for details.

- D0 head D1 flow rate short total * D2 D3 maximum flow rate D4 minimum flow rate D5 temperature D6 maximum temperature D7 minimum temperature velocity * **D**8 D9 nominal target range D10 analog milliamps D11 internal DC volts velocity volts $^{\diamond}$ D12
- D13 auxiliary input volts
- D14 temperature sensor ohms
- D15 self-test checksum
- D16 restarts
- D17 exceptions
- D18 valid echos per 100

security access required

 $^{\diamond}$ applicable to flow calculations requiring velocity sensor



'F' PARAMETER LISTING =

Refer to 'Operation' for details.

F0	enter security code		
F1	emulation mode [*]		
F2	run mode		
F3	keypad to mA output *		
F4	show time		
	set time [*]		
F5	show date		
	set date [*]		
F6	software identification number		
F7	view min/max data		
F8	reset min/max data [*]		
F9	self check [*]		
F10	change security code*		
F11	reset master totalizer*		
F12	force a cold start		
F13	auto zero calibration*		
F14	examine data log:		
	task: view daily totals		
	view flow rates		
	view min/max data		
	method:		
	first day		
	last day		
	specified day		

specified day

F15 clear data log^{*}

* security access required



'P' PARAMETER LISTING

Refer to 'Operation' for details.

	•		
P0	language	•	
		0 = english	
		1 = french	
		2 = italian	
		3 = german	
		4 = spanish	
P1	dimensio	nal units	
		linear	velocity
		0 = centimetres	centimetres per second
		1 = inches	inches per second
		2 = feet	feet per second
		3 = metres	metres per second
P2	temperat	ure units	
		0 = Celcius	
		1 = Fahrenheit	
Р3	primary e	element	
		0 = exponential device	
		1 = BS-3680 Rectangular Flu	me
		2 = BS-3680 Round Nose Ho	rizontal Crest Weir
		3 = BS-3680 Trapezoidal Flui	me
		4 = BS-3680 U-throated Flum	e
		5 = BS-3680 Finite Crest Wei	r
		6 = BS-3680 Thin Plate Recta	angular Weir
		7 = BS-3680 Thin Plate V-not	ch Weir
		8 = Rectangular Weir (contra	cted)
		9 = Round Pipe	
		10 = Palmer Bowlus Flume	
		11 = H Flume	
		12 - Universal Head ve Eleve	

12 = Universal Head vs. Flow

- 13 = Rectangular Area x Velocity
- 14 = Trapezoidal Area x Velocity
- 15 = Modified Trapezoidal Area x Velocity
- 16 = U-channel Area x Velocity
- 17 = Circular Area x Velocity
- 18 = Gull-wing Area x Velocity
- 19 = Egg-shaped Area x Velocity
- 20 = Universal Area x Velocity

P4 method of calculation

- 0 = absolute
- 1 = ratiometric

P5 flow rate units

volume

0 = litres per second lit	tres
1 = cubic feet per second	cubic feet
2 = imperial gallons per minute	imperial gallons
3 = U.S. gallons per minute	U.S. gallons
4 = imperial million gallons per d	lay imperial million gallons
5 = U.S. million gallons per day	U.S. million gallons
6 = cubic metres per hour	cubic metres
7 = cubic metres per day	cubic metres

P6 flow at maximum head

- P7 height of maximum head
- P8 volts in at zero velocity
- P9 velocity at 5 volts in
- P10 velocity at maximum flow

0 = off

1 = low

2 = med

3 = high

P14 display lighting

0 = on 1 = auto off 2 = off

P15 / P18 / P21	relay 1 / 2 / 3 assignment
	0 = not in service
	1 = de-energize on loss of echo
	2 = energize on loss of echo
	3 = de-energize on high flow rate
	4 = energize on high flow rate
	5 = de-energize on low flow rate
	6 = energize on low flow rate
	7 = de-energize on high head
	8 = energize on high head
	9 = de-energize on low head
	10 = energize on low head
	11 = de-energize on high velocity
	12 = energize on high velocity
	13 = de-energize on low velocity
	14 = energize on low velocity
	15 = de-energize on high analog
	16 = energize on high analog
	17 = de-energize on low analog
	18 = energize on low analog
	19 = de-energize on low D11 volts
	20 = energize on low D11 volts
	21 = de-energize on high D11 volts
	22 = energize on high D11 volts
	23 = de-energize on low Aux. volts

	24 = energize on low Aux. volts
	25 = de-energize on high Aux. volts
	26 = energize on high Aux. volts
	27 = de-energize up control on head
	28 = energize up control on head
	29 = de-energize down control on head
	30 = energize down control on head
	31 = de-energize up control on analog
	32 = energize up control on analog
	33 = de-energize down control on analog
	34 = energize down control on analog
	35 = (pulse) flow totalizer
	36 = (pulse) sampler by volume
	37 = (pulse) sampler by time
	38 = (pulse) by time of day
P16 / P19 / P22	relay 1 / 2 / 3 high setpoint
	·····; ··· _· · ····g.· ··· +·····
P17 / P20 / P23	relay 1 / 2 / 3 low setpoint
D 04 • ·	
P24 mA assig	Inment
	0 = flow rate
	1 = head

- 2 = velocity
- 3 = temperature

P25 If custom mA, 20 mA = ?

0 = normal Ø = custom

P26 mA span

0 = 4-20 mA 1 = 0-20 mA P27 mA damping (secs)

P28 mA options (emulator tracking)

0 = don't track emulator

1 = track emulator

P29 fail-safe time (secs)

P30 fail-safe analog mode

0 = hold last value 1 = assume value in P31

P31 fail-safe analog mA (default value)

P32 totalizer multiplier

0 = x 1/1000 (0.001) 1 = x 1/100 (0.01) 2 = x 1/10 (0.1) 3 = x 1 4 = x 10 5 = x 1006 = x 1000

P33 flow rate display (decimal point)

0 = no decimal places 1 = 1 2 = 2 3 = 3 4 = 4 0 = never print

- 1 = interval to be in minutes
- 2 = interval to be in hours
- 3 = print once each day

P35 printer timing

P36 measurement interval

P37 serial data rate

P38 site number

fixed

0 = 1 min	3 = 30
1 = 5	4 = 60
2 = 15	5 = 24 hr

variable (condition)

6 = 15/1 min (% flo	w / min)	19 = 60 / 1	(flow)
7 = 15 / 5	"	20 = 60 / 5	"
8 = 30 / 1	"	21 = 24 hr / 1 min	n
9 = 30 / 5	"	22 = 24 hr / 5 min	п
10 = 60 / 1	"	23 = 24 hr / 15 mi	n "
11 = 60 / 5	"	24 = 15 / 1 min	(head)
12 = 24 hr / 1 min	"	25 = 15 / 5	"
13 = 24 hr / 5 min	п	26 = 30 / 1	"
14 = 24 hr / 15 min	"	27 = 30 / 5	"
15 = 15 / 1 min (f	low)	28 = 60 / 1	"
16 = 15 / 5	"	29 = 60 / 5	"
17 = 30 / 1	"	30 = 24 hr / 1 min	"
18 = 30 / 5	п	31 = 24 hr / 5 min	"
		32 = 24 hr / 15 mi	n "

P40 log rapid setpoint

P39 variable logging condition units

% flow / min	% change of maximum flow per minute
flow	% of maximum flow
head	% of maximum head

P41 log normal setpoint

P39 variable logging condition	units
% flow / min	% change of maximum flow per minute
flow	% of maximum flow
head	% of maximum head

P42 head determination

P43 volts in for zero head

P44 head at 5 volts in

P45 low flow cut-off head

P46 range at zero head

P47 blanking distance

'U' PARAMETERS FOR P3 PRIMARY ELEMENT

The number of 'U' parameters required varies according to the primary element chosen (P3) and the method of calculation (P4). The OCM-3 prompts the user by displaying the next required parameter, insuring the programming is complete.

The following is a list of the specific primary elements to which the OCM-3 can be applied.

Refer to the page covering your particular application; the rest may be disregarded.

P3 primary element

- 0 exponential device (e.g. proportional, V-notch, Parshall etc)
- 1 BS-3680 Rectangular Flume (ISO 4359)
- 2 BS-3680 Round Nose Horizontal Crest Weir (ISO 4374)
- 3 BS-3680 Trapezoidal Flume (ISO 4359)
- 4 BS-3680 U-throated Flume (ISO 4359)
- 5 BS-3680 Finite Crest Weir (ISO 3846)
- 6 BS-3680 Thin Plate Rectangular Weir (ISO 1438/1)
- 7 BS-3680 Thin Plate V Notch Weir (ISO 1438/1)
- 8 Rectangular Weir (contracted)
- 9 Round Pipe
- 10 Palmer-Bowlus Flume
- 11 H Flume
- 12 Universal Head vs. Flow
- 13 Rectangular Area x Velocity
- 14 Trapezoidal Area x Velocity
- 15 Modified Trapezoidal Area x Velocity
- 16 U-channel Area x Velocity

- 17 Circular Area x Velocity
- 18 Gull Wing Area x Velocity
- 19 Egg-shaped Area x Velocity
- 20 Universal Area x Velocity

The primary element must be installed in accordance with the manufacturers recommendations and in accordance with all governing regulations.

'U' parameters required *

U0 = exponent

U1 = k factor (P4 = 0 only)

Typical Exponential Devices:

- » Sutro (proportional) weir
- » head measurement only
- » Rectangular (suppressed) or Trapezoidal (Cipolletti) weir
- » Kahfagi venturi
- » Parshall flume
- » Leopold Lagco
- » Triangular (V-notch) weir

^{*}obtain from manufacturer's specifications.

Reference

```
ABSOLUTE CALCULATION, P4 = 0^{\alpha}
```

For flows that can be calculated by the equation:

 $q = k h^{x}$

where q = flowrate x = exponent (U0)k = constant factor (U1) h = head

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} (h/h_{cal})^{x}$

where	q = flowrate	q _{cal} = flowrate at maximum head
	h = head	h _{cal} = maximum head
	x = exponent (U0)	

Refer to manufacturers specifications for the exponent value.

^a Refer to Operation \ Flow Calculation.

TYPICAL SHARP-CRESTED WEIRS



For rated flows under free flow conditions, the head is measured upstream of the weir plate at a minimum distance of 3 times the maximum head (i.e. where the liquid surface is not affected by drawdown).

SIMPLE EXPONENTIAL DEVICES, P3 = 0

KHAFAGI VENTURI



plan



For rated flows under free flow conditions, the head is measured 15 cm (6") upstream from the beginning of the converging section.

SIMPLE EXPONENTIAL DEVICES, P3 = 0

TYPICAL PARSHALL FLUME



For rated flows under free flow conditions, the head is measured at 2/3 the length of the converging section upstream of the beginning of the throat section.

SIMPLE EXPONENTIAL DEVICES, P3 = 0

TYPICAL LEOPOLD LAGCO



For rated flows under free flow conditions, the head is measured at a point upstream referenced to the beginning of the converging section. Refer to the following table.

Flume Size	Point of Measurement	
(pipe dia. in inches)	mm	inches
4 - 12	25	1.0
15	32	1.3
18	38	1.5
21	44	1.8
24	51	2.1
30	64	2.5
36	76	3.0
42	89	3.5
48	102	4.0
54	114	4.5
60	127	5.0
66	140	5.5
72	152	6.0

BS-3680 Rectangular Flume, P3 = 1

'U' parameters required *	
в	U4 = Cv
b	U5 = Cd
р	U6 = A
L	
	B b p

'U' parameters calculated **

*obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = (2/3)^{1.5} x g^{0.5} x Cv x Cs x Cd x B x h^{1.5}$

where :	q = flow rate	Cv = velocity coefficient
	b = throat width	Cs = shape coefficient
	g = gravitational acceleration	Cd = discharge coefficient
	h = head	

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times Cd/Cd_{cal} \times Cv/Cv_{cal} \times (h/h_{cal})^{1.5}$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ Cv = velocity coefficient $Cv_{cal} = velocity coefficient for maximum head$ Cd = discharge coefficient for head $Cd_{cal} = discharge coefficient for maximum head$

^a Refer to Operation \ Flow Calculation.







* The transducer must be above the maximum head by at least the blanking value, P47.

OCM III

BS-3680 Round Nose Horizontal Crest Weir, P3 = 2

'U' parameters require	ed *	'U' parameters calculated **
U0 = crest width	b	U3 = Cv
U1 = crest height	р	U4 = Cd
U2 = crest length	L	U5 = A

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

$$q = (2/3)^{1.5} x g^{0.5} x Cv x Cs x Cd x b x h^{1.5}$$

where :	q = flow rate	Cv = velocity coefficient
	g = gravitational acceleration	Cs = shape coefficient
	b = throat width	Cd = discharge coefficient
	h = head	

RATIOMETRIC CALCULATION, $P4 = 1^{x}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times Cd/Cd_{cal} \times Cv/Cv_{cal} \times (h/h_{cal})^{1.5}$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ Cv = velocity coefficient $Cv_{cal} = velocity coefficient for maximum head$ Cd = discharge coefficient for maximum head $Cd_{cal} = discharge coefficient for maximum head$

^a Refer to Operation \ Flow Calculation.

BS-3680 ROUND NOSE HORIZONTAL CREST WEIR



BS-3680 TRAPEZOIDAL FLUME, P3 = 3

'U' parameters require	ed *	'U' parameters calculated **
U0 = approach width	в	U5 = Cv
U1 = throat width	b	U6 = Cd
U2 = hump height	р	U7 = Cs
U3 = throat length	L	U8 = A
U4 = slope	m	

*obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

$$q = (2/3)^{1.5} \times g^{0.5} \times Cv \times Cs \times Cd \times b \times h^{1.5}$$

where :	q = flow rate	Cv = velocity coefficient
	g = gravitational acceleration	Cs = shape coefficient
	b = throat width	Cd = discharge coefficient
	h = head	

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times Cs/Cs_{cal} \times Cd/Cd_{al} \times Cv/Cv_{cal} \times (h/h_{cal})^{1.5}$

where :	
q = flow rate	Cs _{cal} = shape coefficient for maximum head
q _{cal} = flow rate at maximum head	Cv = velocity coefficient
h = head	Cvcal = velocity coefficient for maximum head
h _{cal} = maximum head	Cd = discharge coefficient for head
Cs = shape coefficient for head	Cd_{cal} = discharge coefficient for maximum head

^{Refer to Operation \ Flow Calculation.}





* The transducer must be above the maximum head by at least the blanking value, P47.

OCM III

BS-3680 U-Flume, P3 = 4

'U' parameters required	1*	'U' parameters calculated **
U0 = approach diameter	Da	U4 = Cv
U1 = throat diameter	D	U5 = Cd
U2 = hump height	р	U6 = Cu
U3 = throat length	L	U7 = A

^{*}obtain from manufacturer's specifications.

calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = (2/3)^{1.5} x g^{0.5} x Cv x Cu x Cd x D x h^{1.5}$

where :	q = flow rate	Cv = velocity coefficient
	g = gravitational acceleration	Cu = shape coefficient
	b = throat width	Cd = discharge coefficient
	h = head	

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times Cu/Cu_{cal} \times Cd/Cd_{cal} \times Cv/Cv_{cal} \times (h/h_{cal})^{1.5}$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ Cu = shape coefficient for head $Cu_{cal} = shape coefficient for maximum head$ Cv = velocity coefficient $Cv_{cal} = velocity coefficient for maximum head$ Cd = discharge coefficient for head $Cd_{cal} = discharge coefficient for maximum head$

^{α} Refer to Operation \ Flow Calculation.







* The transducer must be above the maximum head by at least the blanking value, P47.

OCM III

BS-3680 FINITE CREST WEIR, P3 = 5

'U' parameters required *		'U' parameters calculated **
U0 = crest width	b	U3 = C
U1 = crest height	р	U4 = Cp
U2 = crest length	L	

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = (2/3)^{1.5} \times g^{0.5} \times C \times Cp \times b \times h^{1.5}$

where :	q = flow rate	C = a function of h and L
	g = gravitational acceleration	
	b = crest width	Cp = a correction factor as a function
	h = head	of h and p applied to C

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times C/C_{cal} \times Cp/Cp_{cal} \times (h/h_{cal})^{1.5}$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ C = discharge coefficient for head $C_{cal} = discharge coefficient for maximum head$ Cp = correction factor for C $Cp_{cal} = correction factor for C_{cal}$

^{α} Refer to Operation \ Flow Calculation.

BS-3680 FINITE CREST WEIR



BS-3680 THIN PLATE RECTANGULAR WEIR, P3 = 6

<i>'U' parameters required *</i>		'U' parameters calculated **
U0 = approach width	В	U3 = Ce
U1 = crest width	b	U4 = K _b
U2 = crest height	р	

*obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = Ce x^{2/3} \sqrt{2q} x be x(h_{e})^{1.5}$

where :	q = flow rate	Ce = a function of h, p, b and B
	g = gravitational acceleration	$be = effective crest width (b + K_b)$
	b = crest width	he = effective head (h + Kh)
	h = head	K _h = 1 mm

RATIOMETRIC CALCULATION. P4 = 1^{α}

For flows that can be calculated by the equation:

 $q = q_{cal} x Ce/Ce_{cal} x (h/h_{cal})^{1.5}$

where : q = flow rate q_{cal} = flow rate at maximum head h = head h_{cal} = maximum head Ce = discharge coefficient for head Cvcal = discharge coefficient for maximum head

^a Refer to Operation \ Flow Calculation.



BS-3680 THIN PLATE V-NOTCH WEIR, P3 = 7

'U' parameters require	d *	'U' parameters calculated **
U0 = notch angle	alpha	U1 = Ce

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = Ce \times 8/15 \times tan(\alpha/2) \times (2g)^{0.5} \times h^{2.5}$

where :	q = flow rate	h = head
	g = gravitational acceleration	Ce = function of h and α
	α = notch angle alpha ^{\diamond}	

^o angle is restricted to 90°, 53.133° or 28.066°. Otherwise use ratiometric calculation.

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times Ce/Ce_{cal} \times (h/h_{cal})^{2.5}$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ Ce = discharge coefficient for head $Ce_{cal} = discharge coefficient for maximum head$

^a Refer to Operation \ Flow Calculation.



RECTANGULAR WEIR (CONTRACTED), P3 = 8

'U' parameters required *

U0 = crest width b

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = K x (b - 0.2h) x h^{1.5}$

where : q = flow rate h = head K = constant

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

 $q = q_{cal} x (b - 0.2h)/(b - 0.2h_{cal}) x (h/h_{cal})^{2.5}$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$

^a Refer to Operation \ Flow Calculation.



* The transducer must be above the maximum head by at least the blanking value, P47.

OCM III
ROUND PIPE, P3 = 9

(based on the Manning Formula)

'U' parameters required *
U0 = pipe inside diameter D
U1 = slope (fall/run) s
U2 = roughness coefficientn

*obtain from manufacturer's specifications.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

where :
$$q = flow rate$$

 $h = head$
 $K = constant$
 $f(h) = A \times R^{0.66}$
 $A = cross sectional area$
 $R = hydraulic radius$
 $s = slope of hydraulic gradient$
 $n = roughness coefficient$

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

-1 -loan		
where :	q = flow rate	h = head
	q _{cal} = flow rate at maximum head	h _{cal} = maximum head

^{$^{\alpha}$} Refer to Operation \ Flow Calculation.

 $q = q_{cal} x f(h)/f(h_{cal})$



- * This dimension should be at least 15 cm (6") shorter than the blanking value, P47.
- ** The transducer must be above the maximum head by at least the blanking value, P47.

PALMER-BOWLUS FLUME *, P3 = 10

'U' parameters required **

U0 = maximum flume width, h_{max}

*typically those manufactured by Warminster or Plasti-Fab.

** obtain from manufacturer's specifications.

Reference

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

 $q = q_{cal} x f(h/h_{max}) / f(h_{cal}/h_{max})$

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ $f(h/h_{max})$ is determined by polynomial synthesis

^a Refer to Operation \ Flow Calculation.

PALMER-BOWLUS FLUME



- * for rated flows under free flow conditions
- ** The transducer must be above the maximum head by at least the blanking value, P47.

H-FLUME *, P3 = 11

'U' parameters required *

U0 = maximum listed head, hmax

^{*}as developed by the U.S. Department of Agriculture, Soil Conservation Service.

** obtain from manufacturer's specifications.

Reference

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} x f(h/h_{max})/f(h_{cal}/h_{max})$

where : q = flow rate $q_{cal} = flow rate at maximum head$ $f(h/h_{max})$ and $f(h_{cal}/h_{max})$ are determined by polynomial synthesis

^a Refer to Operation \ Flow Calculation.



For rated flows under free flow conditions, the head is measured at a point downstream from the flume entrance. Refer to the following table.

Flume Size	Point of Measurement	
D (feet)	cm	inches
0.5	4.7	1.88
0.75	6.7	2.69
1.0	9.1	3.63
1.5	13.5	5.38
2.0	17.9	7.19
2.5	22.5	9.00
3.0	27.2	10.88
4.5	40.5	16.19

* The transducer must be above the maximum head by at least the blanking value, P47.

The flow curve is characterized by entering the head (A_{even}) and flow (A_{odd}) ordinates for the number of data points (n, 4 to 16) over the flow range. The first point (A0,A1) generally being at 0 head and the last point (A_{2n-2} , A_{2n-1}) generally being at maximum head .

'U' parameters required *

U0 = number of data points (n, 4 to 16)

 $A_{even} = head$ $A_{odd} = flow rate$

^{*}obtain from manufacturer's specifications.

Reference

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

 $q = q_{cal} x f(h)/f(h_{cal})$

where : f(h) and f(hcal) are polynomials based on interpolation of data points

^a Refer to Operation \ Flow Calculation.

UNIVERSAL HEAD vs FLOW

e.g. typical compound weir



Parameters

U	0	=	1	1
v	υ.	_		

A0 = 0	head, point 1
A1 = 0	flow, point 1
A2 = 0.3	head, point 2
A3 = 0.1	flow, point 2
A4 = 0.6	head, point 3
A5 = 0.7	flow, point 3
A6 = 0.8	head, point 4
A7 = 1.5	flow, point 4
A8 = 0.95	head, point 5
A9 = 2.25	flow, point 5
A10 = 1	head, point 6
A11 = 2.5	flow, point 6

A12 = 1.05	head, point 7
A13 = 2.65	flow, point 7
A14 = 1.2	head, point 8
A15 = 3.4	flow, point 8
A16 = 1.4	head, point 9
A17 = 5.0	flow, point 9
A18 = 1.7	head, point 10
A19 = 8.0	flow, point 10
A20 = 2	head, point 11
A21 = 11.8	flow, point 11



For best accuracy concentrate data points where changes in flow rate are the greatest.

OCM III

RECTANGULAR AREA X VELOCITY, P3 = 13

'U' parameters required *	'U' parameters calculated **
U0 = channel width B	U1 = area (h)

*obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = 1/1000 \times B \times h \times V$

where : q = flow rate in l/sec B = channel width in cm h = head in cm V = velocity in cm/sec

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

q = q_{cal} x A/A_{cal} x v/v_{cal}

where : q = flow rate $q_{cal} = flow rate at maximum head$ h = head $h_{cal} = maximum head$ v = velocity $v_{cal} = velocity at maximum head$

^a Refer to Operation \ Flow Calculation.



* The transducer must be above the maximum head by at least the blanking value, P47.

TRAPEZOIDAL AREA X VELOCITY, P3 = 14

'U'	parameters	required *
-----	------------	------------

U0 = channel top width B

U3 = area (h)

- U1 = channel base width b
- U2 = channel depth ht

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

q = 1/1000 x (b + mh) x v

m = (B - b)/d

where : q = flow rate, l/sec B = channel top width, cm b = channel bottom width, cm d = depth of channel, cm h = head, cm v = flow velocity, cm/sec

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

- q = q_{cal} x A/A_{cal} x v/v_{cal}
- A = (b + mh) x h

 $A_{cal} = (b + mh_{cal}) x h_{cal}$

- m = (B b)/d
- where : q = flow rate, l/sec B = channel top width, cm b = channel bottom width, cm d = depth of channel, cm h = head, cm v = flow velocity, cm/sec

^{α} Refer to Operation \ Flow Calculation.

TRAPEZOIDAL AREA X VELOCITY



* The transducer must be above the maximum head by at least the blanking value, P47.

MODIFIED TRAPEZOIDAL AREA X VELOCITY, P3 = 15

'U'	parameters lis	sted *
-----	----------------	--------

U0 = channel top width B

U1 = channel base width b

U2 = transition height ht

*obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

q = 1/1000 x (b + mh) x v for $h \le d$

q = 1/1000 x ((b + md)d + (B(h - d))) x v for $h \ge d$

m = (B - b)/d

where :	q = flow rate, l/s	d = height of transition, cm
	B = channel top width, cm	h = head, cm
	b = channel base width, cm	v = flow velocity, cm/sec

RATIOMETRIC CALCULATION, $P4 = 1^{\alpha}$

For flows that can be calculated by the equation:

q = q_{cal} x A/A_{cal} x v/v_{cal}

for $h \le d$: A = (b + mh) x h A_{cal} = (b + mh_{cal}) x h_{cal}

for $h \ge d$: A = (b + md) x d + B(h - d) A_{cal} = (b + md) x d + B(h_{cal} - d)

m = (B - b)/d

where :	q = flow rate	d = depth of channel
	q _{cal} = flow rate at h _{cal}	h = head
	B = channel top width, cm	v = flow velocity, cm/sec
	b = channel base width, cm	

^a Refer to Operation \ Flow Calculation.

'U' parameters calculated **

U3 = area (h)



^{*} The transducer must be above the maximum head by at least the blanking value, P47.

U-CHANNEL AREA X VELOCITY, P3 = 16

'U' parameters required *'U' parameters calculated **U0 = base diameter DU1 = area (h)

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

U-CHANNEL AREA x VELOCITY



^{*} The transducer must be above the maximum head by at least the blanking value, P47.

CIRCULAR AREA X VELOCITY, P3 = 17

'U' parameters listed *		'U' parameters calculated **	
U0 = conduit diameter	ID	U1 = area (h)	

^{*}obtain from manufacturer's specifications.

 ** calculated by OCM-3. May be viewed by accessing 'U' parameter.

CIRCULAR AREA X VELOCITY



- * This dimension should be at least 15 cm (6") shorter than the blanking value, P47.
- ** The transducer must be above the maximum head by at least the blanking value, P47.

GULL-WING AREA X VELOCITY, P3 = 18

'U' parameters required	1*	'U' parameters calculated **
U0 = channel base width	b	U4 = area (h)
U1 = lower angle	alpha (α)	
U2 = upper angle	beta (β)	
U3 = transition height	ht	

^{*}obtain from manufacturer's specifications.

^{**}calculated by OCM-3. May be viewed by accessing 'U' parameter.

GULL WING AREA X VELOCITY



* The transducer must be above the maximum head by at least the blanking value, P47.

EGG-SHAPED AREA X VELOCITY, P3 = 19

'U' parameters listed *		
U0 = top radius	R	
U1 = bottom radius	r	
U2 = axial displacement	d	

'U' parameters calculated **

U3 = area (h)

^{*}obtain from manufacturer's specifications.

** calculated by OCM-3. May be viewed by accessing 'U' parameter.

Reference

ABSOLUTE CALCULATION, $P4 = 0^{\alpha}$

For flows that can be calculated by the equation:

 $q = 1/1000 \times A \times v$

where : q = flow rate in l/sec

v = flow velocity in cm/sec

RATIOMETRIC CALCULATION, $P4 = 1^{m}$

For flows that can be calculated by the equation:

 $q = q_{cal} \times A/A_{cal} \times v/v_{cal}$

where : q = flow rate $q_{cal} = flow rate at h_{cal}$ A = cross sectional area of flow $A_{cal} = cross sectional area of flow at h_{cal}$ v = flow velocity

^a Refer to Operation \ Flow Calculation.

EGG-SHAPED AREA X VELOCITY



- * This dimension should be at least 15 cm (6") shorter than the blanking value, P47.
- ** The transducer must be above the maximum head by at least the blanking value, P47.

UNIVERSAL AREA X VELOCITY, P3 = 20

Arbitrary area is typically used when the primary element and/or its flow do not fit any of the primary elements covered by P3. There is no user difference between absolute and ratiometric calculations P4.

The flow curve is characterized by entering the head (Aeven) and area (Aodd) ordinates for the number of data points (n, 4 to 16) over the flow range. The first point (A0,A1) generally being at 0 head and the last point (A2n-2,A2n-1) generally being at maximum head .

'U' parameters required * 'U' parameters calculated **

U0 = number of data points (n, 4 to 16) U1 = area (h)

Aeven = head

 $A_{odd} = area$

^{*}obtain from manufacturer's specifications.

calculated by OCM-3. May be viewed by accessing 'U' parameter.



* The transducer must be above the maximum head by at least the blanking value, P47.



A even

U0 = 8

A0 = 0	head, point 1
A1 = 0.0	area, point 1
A2 = 0.05	head, point 2
A3 = 0.03	area, point 2
A4 = 0.1	head, point 3
A5 = 0.06	area, point 3
A6 = 0.15	head, point 4
A7 = 0.09	area, point 4
A8 = 0.25	head, point 5
A9 = 0.18	area, point 5
A10 = 0.45	head, point 6
A11 = 0.35	area, point 6
A8 = 0.70	head, point 7
A9 = 0.59	area, point 7
A8 = 1.00	head, point 8
A9 = 0.89	area, point 8

UNIVERSAL AREA X VELOCITY

For best accuracy concentrate data points where changes in flow are the greatest.



MAINTENANCE

The OCM-3 requires very little maintenance due to its solid-state circuitry. However, a program of periodic preventative maintenance should be initiated. This should include regular inspection, general cleaning, overall system performance checks and standard good housekeeping practices.

A periodic inspection of the transducer is recommended, at which time any build-up of material on the face should be removed.

The enclosure should be cleaned using a vacuum cleaner and a clean, dry brush.

For battery replacement refer to Installation / Installing the Memory Back-up Battery.

ERROR CODES

Error Code Listing

1	number underflow
2	number overflow
3	divide error
4	bad argument
5	invalid parameter
6	system is locked
7	head exceeds BS-3680 spec
8	must use ratiometric
9	invalid angle size
10	invalid selection
11	value is view-only
12	characterizer in use
13	feature not available
14	need whole number
15	invalid date
16	invalid time
20	can't zero auxiliary device
21	b must be greater than R-r

COMMUNICATIONS

The OCM-3 provides serial communication either through Milltronics proprietary bipolar current loop or industry standard RS-232. Communication can be used to provide a video or paper printout of OCM activity. It can also be used to provide remote programming and retrieval of the data log, from devices such as computers and PLCs.

When using the bipolar current loop, communication runs of up to 1,500 m can be achieved, as opposed to the limited runs of 15 m using RS-232. By terminating the bipolar current loop with a Milltronics CVCC, the communication format is then converted to RS-232.



- * Milltronics OCM -3 Utilities Software available
- ** provides conversion to RS -232 or 422

Milltronics provides a standard Utilities software package, for convenient communication between an IBM PC compatible computer and the OCM-3. In addition, the user may opt to develop his own custom software program to perform tasks suited to his specific needs.

Protocol

The protocol for the OCM-3 is as follows:

baud rate:	set via P37		
parity:	none		
stop bits:	1		
word length:	8		

The OCM-3 uses a three wire XON/XOFF serial communication link. When the OCM-3's receive buffer is near full, the OCM-3 sends an XOFF signal to tell the sender of the near full condition.

If the sender is also using XON/XOFF, it will respond to the XOFF signal by suspending its transmission.

Similarly, when the OCM-3 receive buffer is near empty, the OCM-3 sends an XON signal to tell the sender that it is safe to resume sending.

The communication baud rate of the OCM-3 and its correspondent must be the same.

Interconnection



Timed Print Out

The OCM-3 can be programmed to periodically print out OCM data (P34/35). A typical print out will have the following format.

Time	hh:mm:ss
Date	dd/mm/yy
Site Number	#
Head	# units
Velocity [*]	# units
Temperature	# units
Flow Rate	# units
Flow Total	# units

^{*}only for applications using velocity input

When the output is directed to a serial printer, a delay is inserted between each line to allow the printer time to print each line before another is sent. In that way, the simplest of printers may be used without need of XON/OFF or hardware handshaking; only a three wire connection is required. The printer must be at least 40 columns wide. The OCM-3 sends text in upper and lower case. The printer must support both upper and lower case or else treat lower case as if it had been sent in upper case.

The printing interval must divide evenly into the selected units of time for printing to occur at regular intervals.

e.g. P34 = 1, printer mode set for interval in minutes

P35 = 7, print interval set for 7 minutes

Printing starts at 7 minutes after the hour and will occur every 7 minutes thereafter until 56 minutes into the hour. The next print out would not occur until 7 minutes into the next hour.

Since the port used for communication is the same as used for normal serial communication, a software switch controls whether or not the display is sent out on the serial port. The switch is controlled by typing 'v' to allow data transmission or 'w' to stop it. The Milltronics printer emulation program does this automatically.

If the OCM-3 is connected to a computer programmed for printer emulation, the computer will receive the print out data and display it on the video terminal (refer to Milltronics Utilities Software).



The Milltronics Utilities Software expands the human interface capabilities of the OCM-3 as compared to that provided by the simple use of the infra-red programmer.

Utilities Software

- remote OCM-3 programming
- remote OCM-3 display
- saving OCM-3 program parameters
- retrieving log data
- creating spread sheet data log file
- creating ASCII text data log file
- emulating serial printer
- graphic flow rate display

- Programmer
- local programming
- local display

The OCM-3 Utilities software provided by Milltronics uses XON/XOFF to ensure communication reliability. The software is written to be executed on an IBM PC compatible computer.

The standard keyboard emulates the OCM-3 programmer. Pressing the keys corresponding to those displayed on the programmer will activate the same function (see Start Up\Keypad). In addition, special functions are provided using the 'w' and 'v' keys found on keyboard.

For additional information, refer to the 'README.DOC' file provided on the disk.





Custom Programming and Third Party Communication Program

Users who wish to write their own software to communicate with the OCM-3 are urged to implement a fully buffered (interrupt driven) serial link incorporating XON/XOFF.

Some users however, will not have the capacity for such an undertaking. Those users may still communicate with the OCM-3 using the less sophisticated communication capabilities provided by such high level languages as Basic, Pascal or C. The only restriction imposed by using a system that does not use XON/XOFF, is that the user must insure that he does not send long streams of commands to the OCM-3 without pausing periodically to allow the OCM-3 time to process them.

Most users who wish to write their own software to communicate with the OCM-3 will do so for a specific purpose. An example would be to create a customized data log. For this, the user will want that only the numeric values be returned, and none of the descriptive information.

To achieve this, the OCM-3 provides a secondary command parser which is accessed when the OCM-3 receives a command enclosed by *('*). When the leading *'*/' is detected by the primary parser, the OCM-3 diverts the message to the secondary parser. The secondary parser remains in control until it receives the trailing *'*/'.

If the trailing '/' is not received within a few seconds, the secondary parser is aborted and control is reverted to the primary parser automatically.

The OCM-3 response message to a command to the secondary parser is a sequence of ASCII characters terminated by an ASCII carriage return (hex 0D) and an ASCII line feed (hex 0A). All system parameters and most other responses are floating point numbers with a maximum of 6 decimal places.

e.g.	typical OCM-3 display value	response message	
	0	0.000000	
	99.123	99.123000	

The following is a list of commands for the secondary parser:

101	iowing is a li			ne secondary parser.
С	ommand			Response
	/a#/	the conte # = 0 thre		where: $a = A, D^*, F, P$ and U
		e.g.	/P5/	the content of P5
*	secondary This may I the short t	command be useful, fo	parser (which r example, i ne protocol f	eters which may be altered through the h bypasses the security parameter F0). in periodically reading and then resetting or altering the value of a D parameter via
	/dn=#/	where: n	•	ter number
		#	is the valu	e stored in the parameter
		e.g.	/d4=50/, tl	nis sets the minimum flow rate to 50 for D4
	/t0/			ilar to that of the 'Flow and Total' t that the multiplier is not displayed
	/t1/	a maxime effect of f e.g. if the	um of 6 de the totalize e OCM-3 d	rned as a floating point number with cimal places, modified to include the er multiplier. isplay shows 00005678 x 100, the 567800.000000
	/t2/	returns th	ne current	24 hour time: hh:mm:ss
	/t3/	returns th	ne current	date: dd/mm/yy
	/f/	returns th	ne current	flow rate

this is similar to the OCM-3 parameter F2 display, however it is not subject to truncation to less than 6 decimal places.

*

returns a code indicating the relay status

code	relay		
	1	2	3
0	0	0	0
1	1	0	0
2	0	1	0
3	1	1	0
4	0	0	1
5	1	0	1
6	0	1	1
7	1	1	1

/l/ start data log down load.

/I1.dd.mm.yyyy/ data log down load - start date/I2.dd.mm.yyyy/ data log down load - end date/I3/ stop data log down load

Modem Communication

Communication with the OCM-3 over an answer back telephone modem is relatively straight forward. For modem communication over conventional telephone lines, a maximum of 1200 or 2400 baud is recommended. Generally, a lower baud rate is more reliable on noisy lines. Modem should be self answering and data checking capability turn on.

The Milltronics Utilities software can be run via modem. In addition the Terminal Emulation program of the Utilities software package can be used to send control sequences directly to the modem.

The remote modem is defined as the modem connected to the OCM-3. The local modem is defined as the modem connected to the computer.

It should be noted that the modem responses listed in the following discussion serve only to typify what is presently common in current field of RS-232 communication.



Considerations

- set the 'cold start' (F12) default baud to 1200 by removing jumper J1 on board A.
- set both modems and the OCM-3 baud rates to 1200.
- set the remote modem to operate in the 'answer back' mode. Refer to modem's user manual.

Communication

- select serial programming link with OCM-3, from the utilities menu. See Milltronics Utilities Software Disk.

The computer screen will display the letters 'v'. These would normally have gone to the OCM-3 to ask it to send its display, but the modem will have returned it to the screen instead.

Establishing Communication

- type the following:

ATD and the telephone number of the remote modem, then press 'ENTER'.

The modem will dial the number you have entered. When the local modem establishes contact with the remote modem, the local modem sends a message to the screen (typical).

CONNECT

- press the 'v' and 'ENTER' keys to turn on the OCM-3's display-to-serial channel.

The OCM-3 is now on line and communication can be carried out. The user may escape to the main menu of the Utilities software and select another program if so desired.

To close the communication, return to the Terminal Emulation program.

Ending Communication

- type W to break OCM-3 communication with the modem*. The computer screen will display a flashing message 'OCM-3 muted'.
- type: +++ (not displayed on the screen) and then press ENTER

This will cause both modems to switch to the 'command' mode. The local modem sends a message to the screen (typical).

οк

- type: ATH and press the ENTER key.

Both modems will hang up. The local modem sends a message to the screen (typical).

ОК

* If the W is not sent, the OCM-3 and its modem will still be in communication. This may result in the OCM-3 modem being unable to answer calls successfully; depending on the type of modem. OCM-3 software, 3.26 or later, averts this problem by automatically breaking communication between the OCM-3 and its modem after 15 minutes.

www.siemens.com/processautomation

Siemens Milltronics Process Instruments Inc. 1954 Technology Drive, P.O. Box 4225 Peterborough, ON, Canada K9J 7B1 Tel: (705) 745-2431 Fax: (705) 741-0466 Email: techpubs.smpi@siemens.com © Siemens Milltronics Process Instruments Inc. 2005 Subject to change without prior notice



Rev. 1.0